

## cFluor™ B690 Anti-Human CD4 (SK3)

| PRODUCT DETAILS         |  |
|-------------------------|--|
| <b>Catalog Number:</b>  | R7-20153 (100 tests)<br>R7-20154 (25 tests)  |
| <b>Reactivity:</b>      | Human  |
| <b>Clone:</b>           | SK3  |
| <b>Format:</b>          | cFluor™ B690   |
| <b>Isotype:</b>         | Mouse IgG1, κ  |
| <b>Volume Per Test:</b> | 5 µL / test  |
| <b>Application:</b>     | Flow cytometry   |
| <b>Formulation:</b>     | Phosphate-buffered solution, pH 7.2, containing 0.09% sodium azide and 0.2% BSA (Origin USA) |
| <b>Storage:</b>         | 2-8°C and protected from light.<br><b>Do not freeze</b>                                      |


### PRODUCT DESCRIPTION

The SK3 monoclonal antibody binds to human CD4, a 59-kDa type I transmembrane glycoprotein in the immunoglobulin superfamily. The CD4 molecule is expressed predominantly on thymocytes and a subpopulation of mature T-helper lymphocytes. It is also present on monocytes at low levels<sup>1,3</sup>. CD4 plays a role in cell-cell interaction by acting as a co-receptor for MHC class II in antigen recognition. CD4 also has been shown to bind glycoprotein 120 on external envelope of HIV<sup>2</sup>. The antibody was conjugated to a fluorophore and purified by affinity chromatography.

### RECOMMENDED USAGE


Each lot of this antibody is quality control tested using flow cytometric analysis. For flow cytometric staining, the suggested use of this reagent is 5 µL per 1 million cells in a staining volume of 100 µL. If whole blood is analyzed, then use 5 µL per 100 µL. It is recommended that users titrate the antibody to obtain the optimal result for their specific application.

PRODUCT DATA



Human peripheral blood was stained with cFluor™ B690 Anti-Human CD4 (clone SK3) (filled histogram) or cFluor™ B690 mouse IgG1, κ isotype control (open histogram). Data shown is gated on lymphocytes.

Spectral signature of cFluor™ B690 from a Cytek® Aurora 5 laser system equipped with 355 nm, 405 nm, 488 nm, 561 nm and 640 nm lasers using CytekAssaySettings.



REFERENCES

1. Evans RL, et al. 1981. Immunol. 78:544
2. Arno A et al. 1999. J. Infect. Dis. 180:56
3. Muench M, et al. 1997. Blood 89:1364

For Research Use Only. Not intended for use in diagnostic procedures.